

Survey: RF1/24
month: March 2024
Participant-No: 6426202
date of issue: 02.03.24



Referenzinstitut für Bioanalytik

Labor Dr. Spranger
Lindberghstr. 9-13
85051 Ingolstadt



Survey directors

Prof. Dr. C. Knabbe
Prof. Dr. Dr. K.P. Kohse
Prof. Dr. M. Neumaier

Head of RfB

Dr. A. Kessler

Bonn, 18. March 2024

Certificate

We confirm that you have participated in the survey for rheumatoid factors in serum.

You have met the requirements of the survey for the following analytes:

Rheumatoid factors (5) | Anti-CCP (U/ml) (1)

This certificate is according to the current version of the RiliBÄK valid until the end of September 2024.



The number in parentheses characterizes the analytical method used.

The assignment of the number to the respective method and/or the respective instrument is to be taken from the total evaluation.



http://www.rfb.bio/certificate?m=6426202&rv_id=RF241&version=0&uid=6F23464B78F44C488D06713ECDC1D74B

RfB
Referenzinstitut
für Bioanalytik

	C	M	R	D/Dmax	T	LL	UL							
Rheumatoid factors [U/ml]	+	5	A	50.3	0.02	50.0	35.0	65.0						
			B	67.1	-0.44	77.2	54.0	101						
Anti-CCP (U/ml)	+	1	A	13.7	0.04	13.4	7.37	19.5						
[U/ml]			B	22.0	-0.59	29.9	16.4	43.4						



Other kits (number):
1-30(1), 1-43(1), 1-141(1), 3-43(1), 5-12(2), 5-23(1), 5-128(1), 5-154(1), 7-43(1), 7-99(1), 9-154(1),



Other kits (number):
1-23(1), 1-30(1), 1-41(1), 2-04(1), 3-58(1),

